


CRYO SYSTEM

FLARE HISTORICAL DATA - Air Lock

[Back to Menu](#)



Data from materials lock purge oxygen monitor.

Boiloff from cryostat was used to purge airlock. Boiloff rate was calculated as 23.4 SCFH gas. Cryostat is at about 6.3 PSIG during this venting.